

## **RECALL OF BLOOD GLUCOSE TEST STRIPS**

**We have been made aware of a fault with some of the testing strips used with TRUEYOU & TRUEYOU mini blood glucose testing machines.**

**If you are using one of these machines and are using test strips with the following batch number, please stop using these strips and return them to the surgery. These strips are giving FALSE low readings.**

**You will need to request some new testing strips for your machine and we will issue these when you return the strips.**

| <b>Product Name</b> | <b>Item Number</b> | <b>Lot number</b> | <b>Exp. Date</b> |
|---------------------|--------------------|-------------------|------------------|
| TRUEyou FAD         | E3I68-82FDL        | HLS1024GBY        | 28-Apr-18        |
| TRUEyou FAD         | E3I68-82FDL        | HLS1032GBY        | 22-Jun-18        |
| TRUEyou FAD         | E3I68-82FDL        | HLS1033GBY        | 23-Jun-18        |
| TRUEyou FAD         | E3I68-82FDL        | HLS1036GBY        | 24-Jun-18        |
| TRUEyou FAD         | E3I68-82FDL        | HLS1037GBY        | 20-Jun-18        |
| TRUEyou FAD         | E3I68-82FDL        | LS1303GBY         | 17-Feb-18        |
| TRUEyou FAD         | E3I68-82FDL        | LS1304GBY         | 17-Feb-18        |
| TRUEyou FAD         | E4GBY70FDL-24      | HLS1037GBY        | 20-Jun-18        |

### **TESTING KITS with affected strips:**

| <b>Product Name</b> | <b>Item Number</b> | <b>Kit Lot number</b> | <b>Exp. Date</b> |
|---------------------|--------------------|-----------------------|------------------|
| TRUEyou FAD         | E4I68-81FDL        | KS1222                | 28-Apr-18        |
| TRUEyou FAD         | E4I68-81FDL        | KS1223                | 28-Apr-18        |
| TRUEyou FAD         | E4I68-81FDL        | KS1570                | 22-Jun-18        |
| TRUEyou FAD         | E4I68-81FDL        | KS1575                | 28-Apr-18        |
| TRUEyou FAD         | E4I68-81FDL        | KS1784                | 22-Jun-18        |
| TRUEyou FAD         | E4I68-81FDL        | KS1785                | 23-May-18        |

**IF YOU HAVE ANY QUERIES, PLEASE SPEAK TO ONE OF THE NURSING TEAM**